

## EU DECLARATION OF CONFORMITY

DOC NAME					PAGES
EU Declaration of Conformity MDR - Navina Smart Control Unit					1(4)
DOC TYPE	DOC NO	VERSION	ENCLOSURE	DATE	STATE
DC	10107	C	Yes	2024-07-01	Approved
AUTHOR		APPROVED BY		CN/PROMO	
Ramzi, Emad		Jørgensen, Toni		PR-85621	
VERIFIED BY				COPY NO	
Raicevic, Andreas					

We,

Wellspect HealthCare  
Aminogatan 1, P.O. Box 14,  
SE-431 21 Mölndal,  
Sweden

being the Manufacturer, registered in the European Union under Single Registration Number (SRN) No. SE-MF-000026271, of the Navina Smart Control unit, including the products listed in the Annex I to this document, with the following characteristics:

- device class IIa, as determined by Rule 12, according to Regulation (EU) 2017/745, Annex VIII
- intended for Transanal irrigation
- GMDN code: 61113
- EMDN category G / code(s):
  - o G99 Gastrointestinal Devices - Other
- Basic UDIDI/Global Model Number: 733338724105D5

Declare under our sole responsibility that the product(s) conform to the requirements of Regulation (EU) 2017/745 of the European Parliament and of the Council, on medical devices, and meet(s) the relevant General Safety and Performance Requirements of Annex I.

All devices are designed, manufactured, tested, and released for sale in accordance with the technical documentation according to Annex II and III of regulation (EU) 2017/745 and the applicable standards.

The conformity assessment procedure was performed following Annex IX of EU Regulation 2017/745.

DOC TYPE	DOC NO	VERSION	DATE	PAGES
DC	10107	C	2024-07-01	2 (4)

Directive 2011/65/EU of the European Parliament and of the Council on the restriction of the use of certain hazardous substances in electrical and electronic equipment, including required amendment 2015/863/EU Compliance for Category 8 (Medical devices and equipment) products (RoHS3 directive)

Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonization of the laws of the Member States relating to the making available on the market of radio equipment (RED directive)

This declaration is made based on the EU Technical Documentation Assessment Certificate MDR 780135 issued by the Notified Body:

BSI Group the Netherlands B.V. (2797)  
Say Building, John M. Keynesplein 9, 1066 EP  
Amsterdam, Netherlands

This declaration of Conformity is approved and signed as dated on first page.

Wellspect HealthCare, Mölndal, Sweden

  
-----  
TONI JØRGENSEN  
VICE PRESIDENT QUALITY ASSURANCE & REGULATORY AFFAIRS

DOC TYPE	DOC NO	VERSION	DATE	PAGES
DC	10107	C	2024-07-01	3 (4)

**ANNEX I**

<b>Article (model) No.*</b>	<b>Product Name, Description</b>
69014	Navina Smart Control Unit

\*Generic article number without the 2-digit suffix specific for a region or country destination when distributing an article.

DOC TYPE	DOC NO	VERSION	DATE	PAGES
DC	10107	C	2024-07-01	4 (4)

## REVISION HISTORY

Document Version	Change note/Description
B	New Basic UDI-DI assigned due to an error in previous Basic UDI-DI identifier.
C	The classification rule, changed from rule 11 to rule 12 due to an error in the previous DoC version.